



STATE OF NEW YORK DEPARTMENT OF HEALTH

Wadsworth Center The Governor Nelson A. Rockefeller Empire State Plaza P.O. BOX 509 Albany, New York 12201-0509

NIRAV R. SHAH, M.D., M.P.H.
Commissioner

SUE KELLY
Executive Deputy Commissioner

March 24, 2011

Page 1 of 14

LAB ID: 11642
MR. DAVID REPPERT
USGS NATIONAL WATER QUALITY LABORATORY
PO BOX 25585
DENVER, CO 80225-0585

Dear Mr. Reppert:

Your laboratory's scores for the Environmental Laboratory Approval Program (ELAP) Non-Potable Water/Solid & Chemical Materials/Air & Emissions Chemistry Proficiency Test are enclosed.

ELAP validates the fitness for use of each sample/analyte used in its PT studies. All analytes passed the assigned value verification and homogeneity evaluations prior to shipment for this study. After the study close date samples/analytes are evaluated again for stability in conjunction with historic pass/fail rates. The evaluation procedures used to establish the assigned value are discussed in the ELAP Certification Manual Item 300. All analytes were determined to be fit for use as PT materials for this study.

The assigned value is the 'made to' value for the analyte. The mean for scoring is either the predicted mean from the FoPT Tables or the participant consensus mean. The acceptance limits are centered on the mean for scoring, or for an unbiased analyte scored with fixed percentage, the assigned value. The acceptance limits are calculated as plus/minus three times the standard deviation or the fixed percentage stated in the report as denoted using the % symbol. In some cases, the acceptance limits are manually adjusted to conform with the FoPT table footnotes 4 through 6. In these cases, the standard deviation cannot be used to set the acceptance limits.

The statistical procedures used to establish study means, standard deviations and acceptance limits are described in the ELAP Certification Manual Item 330.

ELAP does not limit the usage of proficiency test reports by individuals or organizations. ELAP only distributes PT reports to the participant and any organization selected by the participant for distribution.

Sincerely,

Daniel Dickinson
Deputy Director & PT Program Coordinator
Environmental Laboratory Approval Program

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 2 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/ Analysis Date	Mean/ Assigned Value	Standard Deviation/ Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Color									
Color	Color	4056	USGS		12.5	14.5	1.9	5.00 - 25.0	Satisfactory
Analyte Code: 1605			I-1250-85		2/3/2011	14.8		46 passed out of 49 reported results.	
Sample: Non Potable Water Demand									
Chemical Oxygen Demand	mg/L	4001	USGS		110	107	7.73	83.9 - 130	Satisfactory
Analyte Code: 1565			I-3561-85		2/1/2011	112		71 passed out of 74 reported results.	
Organic Carbon, Total	mg/L	4001	SM 18-21		47.2	43.8	2.35	36.8 - 50.9	Satisfactory
Analyte Code: 2040			5310B (00)		1/27/2011	44.1		45 passed out of 46 reported results.	
Sample: Non Potable Water Chlorophenoxy Acid Pesticides (Herbicides)									
2,4-D	ug/L	4018	USGS		1.99	1.99	0.771	0.25 - 4.30	Satisfactory
Analyte Code: 8545			O-2060-01		2/2/2011	2.49		33 passed out of 33 reported results.	
Dicamba	ug/L	4018	USGS		6.89	6.49	1.91	0.847 - 12.2	Satisfactory
Analyte Code: 8595			O-2060-01		2/2/2011	8.47		24 passed out of 24 reported results.	
Sample: Non Potable Water Mercury, Low Level									
Low Level Mercury	ng/L	4058	USGS		55.4	61.3	5.24	45.5 - 77.0	Satisfactory
Analyte Code:			I-4464-01		2/15/2011	61.6		9 passed out of 10 reported results.	
Sample: Non Potable Water Inorganic Nutrients									
Ammonia (as N)	mg/L	4007	USGS		8.02	8.18	0.71	6.05 - 10.3	Satisfactory
Analyte Code: 1515			I-2522-90		2/1/2011	8.21		102 passed out of 106 reported results.	
Nitrate (as N)	mg/L	4007	USGS		4.86	4.90	0.354	3.84 - 5.96	Satisfactory
Analyte Code: 1810			I-2545-90		2/4/2011	4.93		97 passed out of 100 reported results.	
Orthophosphate (as P)	mg/L	4007	USGS		0.572	0.567	0.0569	0.396 - 0.738	Satisfactory
Analyte Code: 1870			I-2601-90		2/1/2011	0.56		80 passed out of 86 reported results.	

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 3 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/ Analysis Date	Mean/ Assigned Value	Standard Deviation/ Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Surfactants									
Surfactants (as MBAS)	mg/L	4043	USGS		0.38	0.514	0.0709	0.302 - 0.727	Satisfactory
Analyte Code: 2025			O-3128-95		1/28/2011	.50		33 passed out of 36 reported results.	
Sample: Non Potable Water Metals I and II									
Silver, Total	ug/L	4011A	USGS		97.4	100	4.93	85.2 - 115	Satisfactory
Analyte Code: 1150			I-4471-97		2/3/2011	100		111 passed out of 118 reported results.	
Silver, Total	ug/L	4011A	USGS		92.0	100	4.93	85.2 - 115	Satisfactory
Analyte Code: 1150			I-4472-97		2/17/2011	100		111 passed out of 118 reported results.	
Aluminum, Total	ug/L	4011A	USGS		1130	1120	69.7	906 - 1320	Satisfactory
Analyte Code: 1000			I-4472-97		2/17/2011	1120		93 passed out of 93 reported results.	
Aluminum, Total	ug/L	4011A	USGS		1020	1120	69.7	906 - 1320	Satisfactory
Analyte Code: 1000			I-4471-97		2/3/2011	1120		93 passed out of 93 reported results.	
Arsenic, Total	ug/L	4011A	USGS		324	342	19.4	284 - 401	Satisfactory
Analyte Code: 1010			I-4020-05		2/17/2011	341		122 passed out of 122 reported results.	
Barium, Total	ug/L	4011A	USGS		1190	1180	51.1	1020 - 1330	Satisfactory
Analyte Code: 1015			I-4471-97		2/3/2011	1180		101 passed out of 102 reported results.	
Barium, Total	ug/L	4011A	USGS		1180	1180	51.1	1020 - 1330	Satisfactory
Analyte Code: 1015			I-4472-97		2/17/2011	1180		101 passed out of 102 reported results.	
Beryllium, Total	ug/L	4011A	USGS		654	673	31.6	579 - 768	Satisfactory
Analyte Code: 1020			I-4471-97		2/3/2011	680		103 passed out of 103 reported results.	
Beryllium, Total	ug/L	4011A	USGS		656	673	31.6	579 - 768	Satisfactory
Analyte Code: 1020			I-4472-97		2/17/2011	680		103 passed out of 103 reported results.	

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 4 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/Analysis Date	Mean/Assigned Value	Standard Deviation/Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Metals I and II									
Cadmium, Total Analyte Code: 1030	ug/L	4011A	USGS I-4471-97		< 0.40 2/3/2011	< 5.90			Satisfactory <i>110 passed out of 126 reported results.</i>
Cadmium, Total Analyte Code: 1030	ug/L	4011A	USGS I-4472-97		< 0.05 2/17/2011	< 5.90			Satisfactory <i>110 passed out of 126 reported results.</i>
Cobalt, Total Analyte Code: 1050	ug/L	4011A	USGS I-4471-97		768 2/3/2011	759 759	30.5	667 - 850	Satisfactory <i>82 passed out of 84 reported results.</i>
Cobalt, Total Analyte Code: 1050	ug/L	4011A	USGS I-4020-05		732 2/17/2011	759 759	30.5	667 - 850	Satisfactory <i>82 passed out of 84 reported results.</i>
Chromium, Total Analyte Code: 1040	ug/L	4011A	USGS I-4020-05		0.247 2/17/2011	< 17.0			Unsatisfactory <== <i>100 passed out of 115 reported results.</i>
Copper, Total Analyte Code: 1055	ug/L	4011A	USGS I-4471-97		587 2/3/2011	623 621	19.6	564 - 682	Satisfactory <i>107 passed out of 120 reported results.</i>
Copper, Total Analyte Code: 1055	ug/L	4011A	USGS I-4020-05		610 2/17/2011	623 621	19.6	564 - 682	Satisfactory <i>107 passed out of 120 reported results.</i>
Iron, Total Analyte Code: 1070	ug/L	4011A	USGS I-4471-97		2110 2/3/2011	2130 2120	84.7	1880 - 2390	Satisfactory <i>79 passed out of 82 reported results.</i>
Manganese, Total Analyte Code: 1090	ug/L	4011A	USGS I-4471-97		2800 2/3/2011	2910 2890	102	2600 - 3210	Satisfactory <i>98 passed out of 105 reported results.</i>
Manganese, Total Analyte Code: 1090	ug/L	4011A	USGS I-4472-97		2770 2/17/2011	2910 2890	102	2600 - 3210	Satisfactory <i>98 passed out of 105 reported results.</i>

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 5 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/ Analysis Date	Mean/ Assigned Value	Standard Deviation/ Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Metals I and II									
Nickel, Total Analyte Code: 1105	ug/L	4011A	USGS I-4020-05		1820 2/17/2011	1930 1910	65.7	1719 - 2130 <i>108 passed out of 116 reported results.</i>	Satisfactory
Nickel, Total Analyte Code: 1105	ug/L	4011A	USGS I-4471-97		1790 2/3/2011	1930 1910	65.7	1719 - 2130 <i>108 passed out of 116 reported results.</i>	Satisfactory
Lead, Total Analyte Code: 1075	ug/L	4011A	USGS I-4472-97		2240 2/17/2011	2390 2400	93	2120 - 2670 <i>137 passed out of 144 reported results.</i>	Satisfactory
Antimony, Total Analyte Code: 1005	ug/L	4011A	USGS I-4472-97		557 2/17/2011	553 580	48.4	407 - 698 <i>103 passed out of 108 reported results.</i>	Satisfactory
Selenium, Total Analyte Code: 1140	ug/L	4011A	USGS I-4020-05		< 0.05 2/17/2011	< 67.0		113 passed out of 116 reported results.	Satisfactory
Strontium, Total Analyte Code: 1160	ug/L	4011A	USGS I-4472-97		219 2/17/2011	221 221	9.78	192 - 251 <i>39 passed out of 42 reported results.</i>	Satisfactory
Thallium, Total Analyte Code: 1165	ug/L	4011A	USGS I-4472-97		314 2/17/2011	319 320	24.5	246 - 393 <i>103 passed out of 105 reported results.</i>	Satisfactory
Vanadium, Total Analyte Code: 1185	ug/L	4011A	USGS I-4020-05		99.0 2/17/2011	99.9 100	4.33	86.9 - 113 <i>80 passed out of 87 reported results.</i>	Satisfactory
Vanadium, Total Analyte Code: 1185	ug/L	4011A	USGS I-4471-97		100 2/3/2011	99.9 100	4.33	86.9 - 113 <i>80 passed out of 87 reported results.</i>	Satisfactory
Zinc, Total Analyte Code: 1190	ug/L	4011A	USGS I-4471-97		350 2/3/2011	384 381	19.3	326 - 441 <i>111 passed out of 114 reported results.</i>	Satisfactory

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 6 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/Analysis Date	Mean/Assigned Value	Standard Deviation/Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Metals I and II									
Zinc, Total	ug/L	4011A	USGS		314	384	19.3	326 - 441	Unsatisfactory <==
Analyte Code: 1190			I-4020-05		2/17/2011	381		111 passed out of 114 reported results.	
Sample: Non Potable Water Metals III									
Molybdenum, Total	ug/L	4011B	USGS		503	518	25.3	442 - 594	Satisfactory
Analyte Code: 1100			I-4472-97		2/17/2011	521		89 passed out of 90 reported results.	
Molybdenum, Total	ug/L	4011B	USGS		500	518	25.3	442 - 594	Satisfactory
Analyte Code: 1100			I-4471-97		2/3/2011	521		89 passed out of 90 reported results.	
Sample: Non Potable Water Minerals									
Chloride	mg/L	4008	USGS		48.6	50.7	2.64	42.8 - 58.6	Satisfactory
Analyte Code: 1575			I-2057-85		1/28/2011	50.4		98 passed out of 100 reported results.	
Sulfate (as SO4)	mg/L	4008	USGS		68.1	69.7	3.79	58.3 - 81.1	Satisfactory
Analyte Code: 2000			I-2057-85		1/28/2011	70.7		74 passed out of 77 reported results.	
Sample: Non Potable Water Minerals II									
Calcium, Total	mg/L	4037	USGS		79.1	83.1	3.22	73.4 - 92.8	Satisfactory
Analyte Code: 1035			I-4471-97		2/3/2011	82.0		69 passed out of 72 reported results.	
Potassium, Total	mg/L	4037	SM 18-21		8.44	8.98	0.622	7.12 - 10.9	Satisfactory
Analyte Code: 1125			3120B (99)		2/3/2011	8.95		63 passed out of 67 reported results.	
Magnesium, Total	mg/L	4037	USGS		30.0	32.4	1.57	27.7 - 37.1	Satisfactory
Analyte Code: 1085			I-4471-97		2/3/2011	32.3		73 passed out of 74 reported results.	
Sodium, Total	mg/L	4037	USGS		11.2	12.4	0.661	10.4 - 14.3	Satisfactory
Analyte Code: 1155			I-4471-97		2/3/2011	12.2		69 passed out of 72 reported results.	

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 7 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/ Analysis Date	Mean/ Assigned Value	Standard Deviation/ Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Nitrite									
Nitrite as N	mg/L	4041	USGS		0.484	0.475	0.0421	0.349 - 0.602	Satisfactory
Analyte Code: 1840			I-2540-90		2/1/2011	0.48		89 passed out of 92 reported results.	
Sample: Non Potable Water Nitrosamines									
N-Nitrosodimethylamine	ug/L	4014	USGS		102	77.7	34.7	16.5 - 181.5	Satisfactory
Analyte Code: 6530			O-3116-87		1/31/2011	165		50 passed out of 50 reported results.	
N-Nitrosodi-n-propylamine	ug/L	4014	USGS		86.3	78.7	16.4	29.6 - 128	Satisfactory
Analyte Code: 6545			O-3116-87		1/31/2011	100		54 passed out of 55 reported results.	
Sample: Non Potable Water Acid extractables									
2,4,6-Trichlorophenol	ug/L	4035	USGS		80.3	91.3	18.1	37.1 - 146	Satisfactory
Analyte Code: 6840			O-3116-87		1/31/2011	116		56 passed out of 57 reported results.	
2,4-Dichlorophenol	ug/L	4035	USGS		118	104	20.1	43.6 - 164	Satisfactory
Analyte Code: 6000			O-3116-87		1/31/2011	134		57 passed out of 57 reported results.	
2,4-Dimethylphenol	ug/L	4035	USGS		< 0.80				Satisfactory
Analyte Code: 6130			O-3116-87		1/31/2011	< 65.0		57 passed out of 57 reported results.	
2,4-Dinitrophenol	ug/L	4035	USGS		< 2.00				Satisfactory
Analyte Code: 6175			O-3116-87		1/31/2011	< 100		56 passed out of 56 reported results.	
2-Chlorophenol	ug/L	4035	USGS		84.5	79.9	16.6	30.1 - 130	Satisfactory
Analyte Code: 5800			O-3116-87		1/31/2011	103		56 passed out of 57 reported results.	
2-Nitrophenol	ug/L	4035	USGS		< 2.12				Satisfactory
Analyte Code: 6490			O-3116-87		1/31/2011	< 50.0		57 passed out of 57 reported results.	
4-Chloro-3-methylphenol	ug/L	4035	USGS		137	137	24.2	64.3 - 209	Satisfactory
Analyte Code: 5700			O-3116-87		1/31/2011	163		56 passed out of 57 reported results.	

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 8 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/ Analysis Date	Mean/ Assigned Value	Standard Deviation/ Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Acid extractables									
Phenol	ug/L	4035	USGS		65.8	86.9	40.2	15.5 - 171	Satisfactory
Analyte Code: 6625			O-3116-87		1/31/2011	155		56 passed out of 57 reported results.	
Sample: Non Potable Water Base/Neutrals									
1,2,4-Trichlorobenzene	ug/L	4036	USGS		< 0.26				Satisfactory
Analyte Code: 5155			O-3116-87		1/31/2011	< 35.0		57 passed out of 57 reported results.	
2,4-Dinitrotoluene	ug/L	4036	USGS		44.1	51.4	9.66	22.4 - 80.3	Satisfactory
Analyte Code: 6185			O-3116-87		1/31/2011	63.1		53 passed out of 55 reported results.	
2,6-Dinitrotoluene	ug/L	4036	USGS		40.7	46.2	8	22.2 - 70.2	Satisfactory
Analyte Code: 6190			O-3116-87		1/31/2011	55.7		54 passed out of 55 reported results.	
2-Chloronaphthalene	ug/L	4036	USGS		< 0.16				Satisfactory
Analyte Code: 5795			O-3116-87		1/31/2011	< 20.0		57 passed out of 57 reported results.	
4-Chlorophenylphenyl ether	ug/L	4036	USGS		27.0	31.3	5.64	14.3 - 48.2	Satisfactory
Analyte Code: 5825			O-3116-87		1/31/2011	37.0		55 passed out of 55 reported results.	
4-Bromophenylphenyl ether	ug/L	4036	USGS		63.6	75.6	15.3	29.8 - 121	Satisfactory
Analyte Code: 5660			O-3116-87		1/31/2011	90.4		54 passed out of 55 reported results.	
Acenaphthene	ug/L	4036	USGS		108	126	21.2	62.6 - 190	Satisfactory
Analyte Code: 5500			O-3116-87		1/31/2011	161		53 passed out of 57 reported results.	
Anthracene	ug/L	4036	USGS		58.3	72.7	12	36.7 - 109	Satisfactory
Analyte Code: 5555			O-3116-87		1/31/2011	86.9		54 passed out of 57 reported results.	
Acenaphthylene	ug/L	4036	USGS		1.29				Unsatisfactory <==
Analyte Code: 5505			O-3116-87		1/31/2011	< 10.0		54 passed out of 57 reported results.	
Bis(2-chloroethyl)ether	ug/L	4036	USGS		152	131	28	47.0 - 215	Satisfactory
Analyte Code: 5765			O-3116-87		1/31/2011	179		52 passed out of 55 reported results.	

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 9 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/ Analysis Date	Mean/ Assigned Value	Standard Deviation/ Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Base/Neutrals									
Bis(2-chloroethoxy)methane Analyte Code: 5760	ug/L	4036	USGS O-3116-87		37.1 1/31/2011	32.1 39.9	5.54	15.5 - 48.8	Satisfactory <i>53 passed out of 55 reported results.</i>
Bis(2-chloroisopropyl)ether Analyte Code: 5780	ug/L	4036	USGS O-3116-87		122 1/31/2011	102 141	22.7	34.0 - 170	Satisfactory <i>54 passed out of 55 reported results.</i>
Bis(2-ethylhexyl)phthalate Analyte Code: 6255	ug/L	4036	USGS O-3116-87		< 2.60 1/31/2011	< 20.0			Satisfactory <i>57 passed out of 58 reported results.</i>
Benzo(a)anthracene Analyte Code: 5575	ug/L	4036	USGS O-3116-87		37.3 1/31/2011	44.2 51.3	7.07	23.0 - 65.5	Satisfactory <i>56 passed out of 57 reported results.</i>
Benzo(a)pyrene Analyte Code: 5580	ug/L	4036	USGS O-3116-87		< 0.32 1/31/2011	< 20.0			Satisfactory <i>56 passed out of 57 reported results.</i>
Benzo(b)fluoranthene Analyte Code: 5585	ug/L	4036	USGS O-3116-87		26.3 1/31/2011	31.9 37.0	6.39	12.7 - 51.1	Satisfactory <i>55 passed out of 57 reported results.</i>
Benzo(g,h,i)perylene Analyte Code: 5590	ug/L	4036	USGS O-3116-87		41.7 1/31/2011	49.7 57.5	9.97	19.8 - 79.6	Satisfactory <i>57 passed out of 57 reported results.</i>
Benzo(k)fluoranthene Analyte Code: 5600	ug/L	4036	USGS O-3116-87		64.9 1/31/2011	73.9 87.5	17.4	21.7 - 126	Satisfactory <i>55 passed out of 57 reported results.</i>
Benzyl butyl phthalate Analyte Code: 5670	ug/L	4036	USGS O-3116-87		< 1.8 1/31/2011	< 50.0			Satisfactory <i>58 passed out of 58 reported results.</i>
Chrysene Analyte Code: 5855	ug/L	4036	USGS O-3116-87		45.5 1/31/2011	49.8 57.6	8.84	23.3 - 76.3	Satisfactory <i>56 passed out of 57 reported results.</i>
Dibenzo(a,h)anthracene Analyte Code: 5895	ug/L	4036	USGS O-3116-87		38.3 1/31/2011	42.5 50.1	9.62	13.7 - 71.4	Satisfactory <i>55 passed out of 57 reported results.</i>

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 10 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/ Analysis Date	Mean/ Assigned Value	Standard Deviation/ Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Base/Neutrals									
Di-n-butyl phthalate Analyte Code: 5925	ug/L	4036	USGS O-3116-87		81.4 1/31/2011	89.5 110	17.9	35.9 - 143	Satisfactory 57 passed out of 58 reported results.
Diethyl phthalate Analyte Code: 6070	ug/L	4036	USGS O-3116-87		85.5 1/31/2011	91.0 117	23.1	21.6 - 160	Satisfactory 57 passed out of 58 reported results.
Dimethyl phthalate Analyte Code: 6135	ug/L	4036	USGS O-3116-87		95.4 1/31/2011	88.1 132	34.1	13.2 - 191	Satisfactory 57 passed out of 58 reported results.
Di-n-octyl phthalate Analyte Code: 6200	ug/L	4036	USGS O-3116-87		101 1/31/2011	125 151	32.1	29.0 - 222	Satisfactory 56 passed out of 58 reported results.
Fluoranthene Analyte Code: 6265	ug/L	4036	USGS O-3116-87		< 0.30 1/31/2011	< 14.0			Satisfactory 55 passed out of 57 reported results.
Hexachlorobenzene Analyte Code: 6275	ug/L	4036	USGS O-3116-87		< 0.30 1/31/2011	< 20.0			Satisfactory 61 passed out of 61 reported results.
Indeno(1,2,3-c,d)pyrene Analyte Code: 6315	ug/L	4036	USGS O-3116-87		54.1 1/31/2011	65.8 84.6	14.1	23.5 - 108	Satisfactory 57 passed out of 57 reported results.
Isophorone Analyte Code: 6320	ug/L	4036	USGS O-3116-87		82.6 1/31/2011	89.1 106	15.9	41.5 - 137	Satisfactory 51 passed out of 55 reported results.
Naphthalene Analyte Code: 5005	ug/L	4036	USGS O-3116-87		51.1 1/31/2011	45.6 60.1	9.39	17.5 - 73.8	Satisfactory 57 passed out of 57 reported results.
Nitrobenzene Analyte Code: 5015	ug/L	4036	USGS O-3116-87		46.5 1/31/2011	42.3 53.8	8.3	17.4 - 67.3	Satisfactory 53 passed out of 55 reported results.
Phenanthrene Analyte Code: 6615	ug/L	4036	USGS O-3116-87		< 0.32 1/31/2011	< 30.0			Satisfactory 56 passed out of 57 reported results.

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 11 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/Analysis Date	Mean/Assigned Value	Standard Deviation/Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Base/Neutrals									
Pyrene	ug/L	4036	USGS		34.5	41.0	8.58	15.2 - 66.7	Satisfactory
Analyte Code: 6665			O-3116-87		1/31/2011	46.7		57 passed out of 57 reported results.	
Sample: Non Potable Water Organic Nutrients									
Kjeldahl Nitrogen, Total	mg/L	4004	USGS		15.9	14.9	1.61	10.1 - 19.8	Satisfactory
Analyte Code: 1795			I-4515-91		2/8/2011	15.3		73 passed out of 78 reported results.	
Phosphorus, Total	mg/L	4004	USGS		2.61	2.52	0.17	2.01 - 3.02	Satisfactory
Analyte Code: 1910			I-4610-91		2/4/2011	2.49		89 passed out of 94 reported results.	
Sample: Non Potable Water Chlorinated Hydrocarbon Pesticides									
4,4-DDE	ug/L	4017	USGS		4.66	5.81	0.947	2.97 - 8.65	Satisfactory
Analyte Code: 7360			O-1126-95		2/16/2011	6.68		46 passed out of 51 reported results.	
alpha-BHC	ug/L	4017	USGS		1.91	1.89	0.41	0.664 - 3.12	Satisfactory
Analyte Code: 7110			O-1126-95		2/16/2011	2.13		51 passed out of 51 reported results.	
Sample: Non Potable Water Purgeables									
1,1,1-Trichloroethane	ug/L	4013	USGS		< 0.75				Satisfactory
Analyte Code: 5160			O-4127-96		2/11/2011	< 10.0		61 passed out of 64 reported results.	
1,1,2,2-Tetrachloroethane	ug/L	4013	USGS		41.6	49.4	7.42	27.2 - 71.7	Satisfactory
Analyte Code: 5110			O-4127-96		2/11/2011	48.1		61 passed out of 63 reported results.	
1,1,2-Trichloroethane	ug/L	4013	USGS		105	108	10.9	75.6 - 141	Satisfactory
Analyte Code: 5165			O-4127-96		2/11/2011	109.3		61 passed out of 63 reported results.	
1,1-Dichloroethane	ug/L	4013	USGS		51.4	53.1	5.9	35.4 - 70.8	Satisfactory
Analyte Code: 4630			O-4127-96		2/11/2011	51.7		62 passed out of 64 reported results.	
1,1-Dichloroethene	ug/L	4013	USGS		70.5	78.0	12.4	40.9 - 115	Satisfactory
Analyte Code: 4640			O-4127-96		2/11/2011	79.5		63 passed out of 64 reported results.	

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 12 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/ Analysis Date	Mean/ Assigned Value	Standard Deviation/ Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Purgeables									
1,2-Dichloroethane Analyte Code: 4635	ug/L	4013	USGS O-4127-96		69.2 2/11/2011	71.2 71.0	7.31	49.3 - 93.2	Satisfactory 62 passed out of 63 reported results.
1,2-Dichlorobenzene Analyte Code: 4610	ug/L	4013	USGS O-4127-96		68.8 2/11/2011	76.2 76.5	7.66	53.2 - 99.2	Satisfactory 72 passed out of 74 reported results.
1,2-Dichloropropane Analyte Code: 4655	ug/L	4013	USGS O-4127-96		33.1 2/11/2011	33.5 33.8	3.89	21.8 - 45.1	Satisfactory 61 passed out of 62 reported results.
1,3-Dichlorobenzene Analyte Code: 4615	ug/L	4013	USGS O-4127-96		47.5 2/11/2011	53.2 54.5	5.46	36.8 - 69.6	Satisfactory 71 passed out of 74 reported results.
1,4-Dichlorobenzene Analyte Code: 4620	ug/L	4013	USGS O-4127-96		30.0 2/11/2011	33.8 34.7	3.52	23.2 - 44.3	Satisfactory 73 passed out of 75 reported results.
Benzene Analyte Code: 4375	ug/L	4013	USGS O-4127-96		54.9 2/11/2011	56.5 56.7	5.19	40.9 - 72.1	Satisfactory 72 passed out of 74 reported results.
Bromodichloromethane Analyte Code: 4395	ug/L	4013	USGS O-4127-96		58.1 2/11/2011	69.3 67.3	7.2	47.7 - 90.9	Satisfactory 58 passed out of 62 reported results.
Bromoform Analyte Code: 4400	ug/L	4013	USGS O-4127-96		< 2.5 2/11/2011	< 5.60			Satisfactory 60 passed out of 62 reported results.
cis-1,3-Dichloropropene Analyte Code: 4680	ug/L	4013	USGS O-4127-96		< 2.5 2/11/2011	< 10.0			Satisfactory 61 passed out of 62 reported results.
Carbon tetrachloride Analyte Code: 4455	ug/L	4013	USGS O-4127-96		93.4 2/11/2011	109 114.7	15.6	62.1 - 156	Satisfactory 61 passed out of 62 reported results.
Trichlorofluoromethane Analyte Code: 5175	ug/L	4013	USGS O-4127-96		< 1.5 2/11/2011	< 8.00			Satisfactory 60 passed out of 61 reported results.

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 13 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 PO BOX 25585
 EPA Lab Code: DENVER, CO 80225-0585
 CO00009 (303) 236-2000
 Director: MR. DAVID REPERT

Shipment Date : 18-Jan-2011
 Closing Date : 03-Mar-2011
 Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/ Analysis Date	Mean/ Assigned Value	Standard Deviation/ Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Purgeables									
Chlorobenzene Analyte Code: 4475	ug/L	4013	USGS O-4127-96		93.5 2/11/2011	92.7 94.0	8.27	67.8 - 117	Satisfactory 68 passed out of 72 reported results.
Dibromochloromethane Analyte Code: 4575	ug/L	4013	USGS O-4127-96		67.1 2/11/2011	80.9 80.4	8.61	55.1 - 107	Satisfactory 61 passed out of 62 reported results.
Chloroethane Analyte Code: 4485	ug/L	4013	USGS O-4127-96		56.9 2/11/2011	62.2 62.2	60%	24.9 - 99.5	Satisfactory 58 passed out of 63 reported results.
Chloroform Analyte Code: 4505	ug/L	4013	USGS O-4127-96		13.5 2/11/2011	13.8 13.4	1.56	9.13 - 18.5	Satisfactory 59 passed out of 63 reported results.
Ethyl benzene Analyte Code: 4765	ug/L	4013	USGS O-4127-96		< 0.9 2/11/2011	< 9.00			Satisfactory 70 passed out of 74 reported results.
Methylene chloride Analyte Code: 4975	ug/L	4013	USGS O-4127-96		81.4 2/11/2011	80.4 80.4	10.4	49.3 - 111	Satisfactory 63 passed out of 64 reported results.
trans-1,3-Dichloropropene Analyte Code: 4685	ug/L	4013	USGS O-4127-96		< 3.5 2/11/2011	< 8.00			Satisfactory 60 passed out of 62 reported results.
Tetrachloroethene Analyte Code: 5115	ug/L	4013	USGS O-4127-96		< 0.65 2/11/2011	< 10.0			Satisfactory 65 passed out of 66 reported results.
Toluene Analyte Code: 5140	ug/L	4013	USGS O-4127-96		< 0.45 2/11/2011	< 7.00			Satisfactory 71 passed out of 75 reported results.
Trichloroethene Analyte Code: 5170	ug/L	4013	USGS O-4127-96		21.7 2/11/2011	21.4 21.7	2.58	13.7 - 29.2	Satisfactory 62 passed out of 67 reported results.
Total Xylenes Analyte Code: 5260	ug/L	4013	USGS O-4127-96		150 2/11/2011	150 156.4	20	89.7 - 210	Satisfactory 72 passed out of 73 reported results.

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 14 of 14

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
PO BOX 25585
EPA Lab Code: DENVER, CO 80225-0585
CO00009 (303) 236-2000
Director: MR. DAVID REPPERT

Shipment Date : 18-Jan-2011
Closing Date : 03-Mar-2011
Score Date : 23-Mar-2011

ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 340 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Test Method	Prep Method	Result/ Analysis Date	Mean/ Assigned Value	Standard Deviation/ Fixed %	Acceptance Limits	Score
Sample: Non Potable Water Purgeables									
Vinyl chloride	ug/L	4013	USGS		< 1.5				Satisfactory
Analyte Code: 5235			O-4127-96		2/11/2011	< 20.0		63 passed out of 65 reported results.	
Sample: Non Potable Water Residue									
Solids, Total	mg/L	4002	USGS		346	358	13.5	318 - 399	Satisfactory
Analyte Code: 1950			I-3750-85		2/9/2011	361		65 passed out of 77 reported results.	
Sample: Non Potable Water Total Dissolved Solids									
Solids, Total Dissolved	mg/L	4040	USGS		182	194	17.676	141 - 247	Satisfactory
Analyte Code: 1955			I-1750-85		2/9/2011	190		71 passed out of 75 reported results.	